

Certificate of Analysis**Residual *E. coli* DNA Quantitation Kit (2G)**Product No.: **1101107-1**Lot No.: **20250901**Date of Mfg.: **03 Sep. 2025**Expiration Date: **07 Apr. 2027**

Test Item	Specification	Result
Linearity of Standard curve	$R^2 \geq 0.990$.	PASS
Amplification efficiency of Standard curve	$E = 90.0\% - 110.0\%$.	PASS
Repeatability	30 fg/ μ L <i>E. coli</i> DNA Control was detected, $CV \leq 20.0\%$.	PASS
No Template Control (NTC)	The Ct value of NTC should be higher than 35.00 or absence of specific amplification.	PASS

Approved By Zong WeyangDAY/MONTH/YEAR 05/09/2025
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China