

Document No.: HEK-BRP-ZL-07

Version No.: A/3

Residual HEK293 DNA Quantitation Kit (3G)Product No.: **1101104-1**Lot No.: **20251101**Date of Mfg.: **14 Nov. 2025**Expiration Date: **06 Mar. 2027**

Test Item	Specification	Result
Linearity of Standard curve	$R^2 \geq 0.990$.	PASS
Amplification efficiency of Standard curve	$E = 90.0\% - 110.0\%$.	PASS
Repeatability	30 fg/ μ L HEK293 DNA Control was detected, $CV \leq 15.0\%$.	PASS
No Template Control (NTC)	The mean value of NTC should not be more than 6.00 fg/ μ L.	PASS
HEK293 DNA Control	Report the results.	31.6 ng/ μ L

Approved By Zong WeijingDAY/MONTH/YEAR 25/11/2025
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China