

Document No.: DNA-BRP-ZL-105

Version No.: A/0

Residual Host Cell DNA Sample Preparation Kit(2G)Product No.: **1104198**Lot No.: **20251101**Date of Mfg.: **17 Nov. 2025**Expiration Date: **31 Jan. 2027**

Test Item	Specification	Result
Recovery	The recovery of samples was between 70.0% -130.0%.	PASS
Negative Control Sample (NCS)	The mean Ct value of NCS should be higher than the mean Ct value of the lowest concentration in the standard curve.	PASS

Approved By Zong WeiyangDAY/MONTH/YEAR 03/12/2025
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China