

Document No.: Human-BRP-ZL-15

Version No.: A/0

Residual Human DNA Quantitation Kit (2G)

Product No.: 1101108

Lot No.: 20250901

Date of Mfg.: 30 Sep. 2025

Expiration Date: 19 Aug. 2027

Test Item	Specification	Result
Linearity of Standard curve	$R^2 \geq 0.990$.	PASS
Amplification efficiency of Standard curve	$E = 90.0\% - 110.0\%$.	PASS
Repeatability	30 fg/ μ L Human DNA Control was detected, $CV \leq 15.0\%$.	PASS
No Template Control (NTC)	The mean value of NTC should not be more than 6.00 fg/ μ L.	PASS
Human DNA Control	Report the results.	30.5 ng/ μ L

Approved By Zong WeiyangDAY/MONTH/YEAR 04/11/2025
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China