

Document No.: NS0-BRP-ZL-04

Version No.: A/2

Residual NS0&SP2/0 DNA Quantitation KitProduct No.: **SK030208N100**Lot No.: **20250902**Date of Mfg.: **17 Sep. 2025**Expiration Date: **13 Jan. 2027**

Test Item	Specification	Result
Linearity of Standard curve	$R^2 \geq 0.990$.	PASS
Amplification efficiency of Standard curve	$E = 90.0\% - 110.0\%$.	PASS
Repeatability	3 fg/ μ L NS0&SP2/0 DNA Control was detected, $CV \leq 20.0\%$.	PASS
No Template Control (NTC)	The Ct value of NTC should not be less than 35.00 or absence of specific amplification.	PASS
NS0&SP2/0 DNA Control	Report the results.	30.0 ng/ μ L

Approved By Zong WeiqingDAY/MONTH/YEAR 26/09/2025
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China