

Document No.: Reverse Transcriptase-BRP-ZL-02

Version No.: A/1

**Reverse Transcriptase Assay Control**

Product No.: 1505701

Lot No.: 20251001

Date of Mfg.: 14 Oct. 2025

Expiration Date: 30 Sep. 2026

Test Item	Specification	Result
Positive Control Sample (PCS)	The Ct value of PCS should be not more than 28.00.	PASS
Negative Control Sample (NCS)	The Ct value of NCS (Buffer A) should be not less than 40.00.	PASS

Approved By Zhang LipingDAY/MONTH/YEAR 15/10/2025  
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access [www.shentekbio.com](http://www.shentekbio.com)For any other inquiries, please contact us by email: [info@shentekbio.com](mailto:info@shentekbio.com)

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China