

Document No.: Sf9 HCP-BRP-ZL-11

Version No.: A/2

**Sf9 HCP ELISA Kit
(One-step ELISA)**Product No.: **1301312**Lot No.: **20250901**Date of Mfg.: **17 Sep. 2025**Expiration Date: **10 Jun. 2026**

Test Item	Specification	Result
Sf9 HCP Calibration Standard	Report the results.	0.82 µg/bottle
Linearity of Standard Curve	$R^2 \geq 0.990$.	PASS
Repeatability	81 ng/mL sample solution was tested, $CV \leq 20.0\%$.	PASS
Limit of Quantification	3 ng/mL sample solution was tested, $CV \leq 25.0\%$, relative bias of concentration was not more than 25.0%.	PASS
Negative Control Sample (NCS)	The mean value of NCS ($OD_{450 \text{ nm}}$) should not be more than 0.200.	PASS

Approved By Cheng JiajunDAY/MONTH/YEAR 18/09/2025
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China