

Document No.: Virus-BRP-ZL-47

Version No.: A/0

**Virus DNA & RNA Extraction Kit**Product No.: **1506730**Lot No.: **20251001**Date of Mfg.: **29 Oct. 2025**Expiration Date: **31 Jan. 2027**

Test Item	Specification	Result
Recovery	The recovery of samples was between 50.0% and 150.0%.	PASS
	The recovery of samples (with matrix) was between 50.0% and 150.0%.	PASS
Negative Control Sample (NCS)	The Ct value of NCS should be larger than the mean Ct value of the lowest concentration in the standard curve.	PASS
No Template Control (NTC)	The Ct value of NTC should not be less than 35.00 cycles.	PASS

Approved By Zhang LipingDAY/MONTH/YEAR 03/11/2025  
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access [www.shentekbio.com](http://www.shentekbio.com)For any other inquiries, please contact us by email: [info@shentekbio.com](mailto:info@shentekbio.com)

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China