

Document No.: HCP-BRP-ZL-19

Version No.: A/1

***E. coli* (Protein Expression Strains) HCP ELISA Kit**Product No.: **1301301**Lot No.: **20250502**Date of Mfg.: **29 May 2025**Expiration Date: **11 Mar. 2026**

Test Item	Specification	Result
Linearity of Standard Curve	$R^2 \geq 0.990$ .	PASS
Repeatability	81 ng/mL sample solution was tested, $CV \leq 15.0\%$ .	PASS
Limit of Quantification	1.5 ng/mL sample solution was tested, $CV \leq 20.0\%$ , relative bias of concentration was not more than 20.0%.	PASS
Negative Control Sample (NCS)	The mean value of NCS ( $OD_{450\text{ nm}}$ ) should be not more than 0.150.	PASS

Approved By Zong WeiyiDAY/MONTH/YEAR 03/06/2025  
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access [www.shentekbio.com](http://www.shentekbio.com)For any other inquiries, please contact us by email: [info@shentekbio.com](mailto:info@shentekbio.com)

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China