

Document No.: 293T HCP-BRP-ZL-08

Version No.: A/2

HEK293 HCP ELISA Kit
(One-step ELISA)

Product No.: 1301311

Lot No.: 20250401

Date of Mfg.: 01 Apr. 2025

Expiration Date: 22 Oct. 2025

Test Item	Specification	Result
Linearity of Standard Curve	$R^2 \geq 0.990$.	PASS
Repeatability	135 ng/mL sample solution was tested, $CV \leq 15.0\%$.	PASS
Limit of Quantification	6 ng/mL sample solution was tested, $CV \leq 20.0\%$, relative bias of concentration should not be more than 20.0%.	PASS
Negative Control Sample (NCS)	The mean value of NCS ($OD_{450\text{ nm}}$) should not be more than 0.200.	PASS

Approved By Zong WeipingDAY/MONTH/YEAR 09/04/2025
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China