

Document No.: MVM-BRP-ZL-05

Version No.: A/0

MVM Detection Kit

Product No.: 1506742

Lot No.: 20250301

Date of Mfg.: 21 Mar. 2025

Expiration Date: 11 Nov. 2026

Test Item	Specification	Result
Low concentration sample	FAM channel: Ct < 40.00 with specific amplification. VIC channel: Ct < 35.00 with specific amplification.	PASS
No Template Control (NTC)	FAM channel: Undetected or absence of specific amplification. VIC channel: Ct < 35.00 with specific amplification.	PASS

Approved By Zong WeiyiDAY/MONTH/YEAR 27/03/2025
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China