

Document No.: Virus-BRP-ZL-49

Version No.: A/0

MVM Positive ControlProduct No.: **1506747**Lot No.: **20250302**Date of Mfg.: **26 Mar. 2025**Expiration Date: **27 Feb. 2026**

Test Item	Specification	Result
Positive Control Sample (PCS)	FAM channel: Ct < 35.00 with specific amplification. VIC channel: Ct < 35.00 with specific amplification.	PASS
Negative Control Sample (NCS)	FAM channel: Undetected or absence of specific amplification. VIC channel: Ct < 35.00 with specific amplification.	PASS

Approved By Zong WeiyangDAY/MONTH/YEAR 28/03/2025
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China