

Document No.: MB-BRP-ZL-10

Version No.: A/2

Mycobacteria DNA Extraction Kit

Product No.: 1503601

Lot No.: 20250301

Date of Mfg.: 21 Mar. 2025

Expiration Date: 31 Jul. 2026

Test Item	Specification	Result
Positive Sample (with matrix)	Mycobacteria DNA in test sample (with matrix) was extracted and analyzed. Ct value of FAM channel<40.00, and Ct value of VIC channel<35.00.	PASS
Positive Sample (without matrix)	Mycobacteria DNA in test sample (without matrix) was extracted and analyzed. Ct value of FAM channel<40.00, and Ct value of VIC channel<35.00.	PASS
Negative Control Sample	Ct value of FAM channel $\geq$ 40.00 or absence of specific amplification, and Ct value of VIC channel<35.00.	PASS

Approved By Zong WeiyangDAY/MONTH/YEAR 27/03/2025
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China