

Document No.: Mycop-BRP-ZL-26

Version No.: A/2

Mycoplasma DNA Detection Kit (2G)Product No.: **1509841**Lot No.: **20250301**Date of Mfg.: **31 Mar. 2025**Expiration Date: **20 Dec. 2025**

Test Item	Specification	Result
Positive Control	Ct value of FAM channel < 35.00, and Ct value of VIC channel < 35.00.	PASS
No Template Control (NTC)	Ct value of FAM channel ≥ 40.00 or absence of specific amplification, and the Ct value of VIC channel < 35.00.	PASS

Approved By Zong WeiyangDAY/MONTH/YEAR 08/04/2025
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China