

Document No.: Mycop-BRP-ZL-29

Version No.: A/2

Mycoplasma DNA Extraction Kit (2G)

Product No.: 1509840

Lot No.: 20250301

Date of Mfg.: 19 Mar. 2025

Expiration Date: 08 Jan. 2026

Test Item	Specification	Result
Positive Sample (with matrix)	Mycoplasma DNA in test sample (with matrix) was extracted and analyzed. Ct value of FAM channel<40.00, and Ct value of VIC channel<35.00.	PASS
Positive Sample (without matrix)	Mycoplasma DNA in test sample (without matrix) was extracted and analyzed. Ct value of FAM channel<40.00, and Ct value of VIC channel<35.00.	PASS
Negative Control Sample	Mycoplasma DNA of Negative control sample was extracted and analyzed. Ct value of FAM channel \geq 40.00 or absence of specific amplification, and Ct value of VIC channel<35.00.	PASS

Approved By Zong WeiqingDAY/MONTH/YEAR 27/03/2025
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China