

Document No.: Mycop-BRP-ZL-29

Version No.: A/2

**Mycoplasma DNA Extraction Kit (2G)**

Product No.: 1509840

Lot No.: 20250401

Date of Mfg.: 27 Apr. 2025

Expiration Date: 17 Dec. 2026

Test Item	Specification	Result
Positive Sample (with matrix)	Mycoplasma DNA in test sample (with matrix) was extracted and analyzed. Ct value of FAM channel<40.00, and Ct value of VIC channel<35.00.	PASS
Positive Sample (without matrix)	Mycoplasma DNA in test sample (without matrix) was extracted and analyzed. Ct value of FAM channel<40.00, and Ct value of VIC channel<35.00.	PASS
Negative Control Sample	Mycoplasma DNA of Negative control sample was extracted and analyzed. Ct value of FAM channel $\geq$ 40.00 or absence of specific amplification, and Ct value of VIC channel<35.00.	PASS

Approved By Zong WeijiangDAY/MONTH/YEAR 08/05/2025  
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access [www.shentekbio.com](http://www.shentekbio.com)For any other inquiries, please contact us by email: [info@shentekbio.com](mailto:info@shentekbio.com)

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China