

Document No.: *P.pastoris* HCP-BRP-ZL-06

Version No.: A/0

***P.pastoris* HCP ELISA Kit**  
**(One-step ELISA)**

Product No.: 1301313

Lot No.: 20250301

Date of Mfg.: 13 Mar. 2025

Expiration Date: 03 Jan. 2026

Test Item	Specification	Result
Linearity of Standard curve	$R^2 \geq 0.990$ .	PASS
Repeatability	50 ng/mL sample solution was tested, $CV \leq 20.0\%$ .	PASS
Limit of Quantification	2 ng/mL sample solution was tested, $CV \leq 25.0\%$ , relative bias of concentration was not more than 25.0%.	PASS
Negative Control Sample (NCS)	The mean value of NCS ( $OD_{450\text{ nm}}$ ) should be not more than 0.300.	PASS

Approved By Zong WeijiangDAY/MONTH/YEAR 25/03/2025  
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access [www.shentekbio.com](http://www.shentekbio.com)For any other inquiries, please contact us by email: [info@shentekbio.com](mailto:info@shentekbio.com)

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China