

Document No.: *P.pastoris* HCP-BRP-ZL-06

Version No.: A/1

***P.pastoris* HCP ELISA Kit**
(One-step ELISA)

Product No.: 1301313

Lot No.: 20250402

Date of Mfg.: 27 Apr. 2025

Expiration Date: 26 Mar. 2026

Test Item	Specification	Result
Linearity of Standard curve	$R^2 \geq 0.990$.	PASS
Repeatability	50 ng/mL sample solution was tested, $CV \leq 20.0\%$.	PASS
Limit of Quantification	2 ng/mL sample solution was tested, $CV \leq 25.0\%$, relative bias of concentration was not more than 25.0%.	PASS
Negative Control Sample (NCS)	The mean value of NCS ($OD_{450\text{ nm}}$) should not be more than 0.300.	PASS

Approved By Zong WeiqingDAY/MONTH/YEAR 09/05/2025
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China