

Document No.: Virus-BRP-ZL-54

Version No.: A/0

REO-3 Detection Kit

Product No.: **1506733**Lot No.: **20250701**Date of Mfg.: **07 Jul. 2025**Expiration Date: **26 Feb. 2027**

| Test Item | Specification | Result |
|---------------------------|---|--------|
| Low concentration sample | FAM channel: Ct < 38.00 and specific amplification. VIC channel: Ct < 35.00 and specific amplification. | PASS |
| No Template Control (NTC) | FAM channel: Ct \geq 40.00 or absence of specific amplification. VIC channel: Ct < 35.00 and specific amplification. | PASS |

Approved By Zong WeiyingDAY/MONTH/YEAR 09/07/2025
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China