

Document No.: Virus-BRP-ZL-53

Version No.: A/0

REO-3 Positive ControlProduct No.: **1506737**Lot No.: **20250701**Date of Mfg.: **07 Jul. 2025**Expiration Date: **21 May 2026**

Test Item	Specification	Result
Positive Control Sample (PCS)	FAM channel: Ct < 35.00 and specific amplification. VIC channel: Ct < 35.00 and specific amplification.	PASS
Negative Control Sample (NCS)	FAM channel: Ct \geq 40.00 or absence of specific amplification. VIC channel: Ct < 35.00 and specific amplification.	PASS

Approved By Zong WeiyiDAY/MONTH/YEAR 09/07/2025
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China