

Document No.: RCL-BRP-ZL-03

Version No.: A/1

Replication-Competent Lentivirus (RCL) Quantitation Kit

Product No.: 1403441

Lot No.: 20250701

Date of Mfg.: 22 Jul. 2025

Expiration Date: 23 Dec. 2026

Test Item	Specification	Result
Linearity of Standard curve	$R^2 \geq 0.990$.	PASS
Amplification efficiency of Standard curve	$E=83.3\%-110.0\%$.	PASS
Repeatability	Low concentration sample was detected, $CV \leq 20.0\%$.	PASS
No Template Control (NTC)	The Ct value of NTC should not be less than 35.00 cycles.	PASS

Approved By Zong WeiyiDAY/MONTH/YEAR 29/07/2025
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China