

Document No.: RCR-BRP-ZL-04

Version No.: A/0

Replication-Competent Retrovirus (RCR) Quantitation KitProduct No.: **1403442**Lot No.: **20250701**Date of Mfg.: **22 Jul. 2025**Expiration Date: **09 Jan. 2027**

Test Item	Specification	Result
Linearity of Standard curve	$R^2 \geq 0.990$.	PASS
Amplification efficiency of Standard curve	$E=83.3\%-110.0\%$.	PASS
Repeatability	Low concentration sample was detected, $CV \leq 20.0\%$.	PASS
No Template Control (NTC)	The Ct value of NTC should not be less than 35.00 or undetermined.	PASS

Approved By Zong WeiyingDAY/MONTH/YEAR 29/07/2025
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China