

Document No.: CHO-BRP-ZL-14

Version No.: A/2

Residual CHO DNA Quantitation Kit (2G)Product No.: **1101100-1**Lot No.: **20250701**Date of Mfg.: **09 Jul. 2025**Expiration Date: **26 May 2027**

Test Item	Specification	Result
Linearity of Standard curve	$R^2 \geq 0.990$.	PASS
Amplification efficiency of Standard curve	$E=90.0\%-110.0\%$.	PASS
Repeatability	3 fg/ μ L CHO DNA Control was detected, $CV \leq 15.0\%$.	PASS
No Template Control (NTC)	The Ct value of NTC should be higher than 35.00 or absence of specific amplification.	PASS

Approved By Zeng WeiyiDAY/MONTH/YEAR 09/07/2025
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China