

Document No.: CHO-BRP-ZL-17

Version No.: A/0

Residual CHO DNA Size Analysis Kit (2G)

Product No.: 1103170-2

Lot No.: 20250401

Date of Mfg.: 11 Apr. 2025

Expiration Date: 30 Dec. 2026

Test Item	Specification		Result
Linearity of Standard curve	R ² ≥0.990.		PASS
Amplification efficiency of Standard curve	E=83.3%-110.0%.		PASS
Repeatability	CHO Primer & Probe MIX-95	300 fg/μL CHO DNA Control was detected, CV≤20.0%.	PASS
	CHO Primer & Probe MIX-110		PASS
	CHO Primer & Probe MIX-215		PASS
	CHO Primer & Probe MIX-523		PASS
No Template Control (NTC)	The mean Ct value of NTC should be higher than the mean Ct value of the lowest concentration in the standard curve or absence of specific amplification.		PASS

Approved By Zong WeiyiDAY/MONTH/YEAR 21/04/2025
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China