

Document No.: *E. coli* RNA-BRP-ZL-08

Version No.: A/2

Residual *E. coli* RNA Quantitation Kit (2G)Product No.: **1201201-1**Lot No.: **20250801**Date of Mfg.: **20 Aug. 2025**Expiration Date: **01 Apr. 2027**

Test Item	Specification	Result
Linearity of Standard curve	$R^2 \geq 0.990$.	PASS
Amplification efficiency of Standard curve	$E=83.3\%-110.0\%$.	PASS
Repeatability	2 fg/ μ L <i>E. coli</i> RNA Control was detected, $CV \leq 15.0\%$.	PASS
No Template Control (NTC)	The average Ct value of NTC should be at least 2 cycles higher than the lowest standard curve.	PASS

Approved By Zong MeiyiDAY/MONTH/YEAR 22/08/2025
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China