

Certificate of Analysis

Document No.: DNA-BRP-ZL-05

Version No.: A/4

Residual Host Cell DNA Sample Preparation Kit

Product No.: **SK030203D100**

Lot No.: 20250201

Date of Mfg.: **06 Feb. 2025**

Expiration Date: 18 Mar. 2026

Test Item	Specification	Result
Recovery	The recovery of samples was between 70.0% -130.0%.	PASS
Extracted Positive Control	The recovery of Positive control was between 80.0%-120.0%.	PASS
Negative Control Sample (NCS)	The Ct value of NCS should be not less than the $Ct_{st6}+2$.	PASS

Approved By Zong Weigag

Certificates available online, please access www.shentekbio.com

For any other inquiries, please contact us by email: info@shentekbio.com

Address:8th Floor,6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China