

Document No.: DNA-BRP-ZL-24

Version No.: A/2

Residual Host Cell DNA Sample Preparation KitProduct No.: **1104191**Lot No.: **20250401**Date of Mfg.: **16 Apr. 2025**Expiration Date: **16 Oct. 2026**

Test Item	Specification	Result
Recovery	The recovery of samples was between 70.0% and 130.0%.	PASS
Positive Control Sample (PCS)	The recovery of PCS was between 80.0% and 120.0%.	PASS
Negative Control Sample (NCS)	The Ct value of NCS should be higher than the mean Ct value of the lowest concentration in the standard curve.	PASS

Approved By Zong WeipingDAY/MONTH/YEAR 23/04/2025
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China