

Document No.: HPV18-BRP-ZL-04

Version No.: A/0

Residual HPV18 E6/E7 DNA Size Analysis Kit

Product No.: SK030307S-P

Lot No.: 20250801

Date of Mfg.: 07 Aug. 2025

Expiration Date: 31 Jul. 2027

Test Item	Specification		Result
Linearity of Standard curve	R ² ≥0.990.		PASS
Amplification efficiency of Standard curve	E=83.3%-110.0%.		PASS
Repeatability	E6-100 primer&probe MIX	Low concentration sample was detected, CV≤30.0%.	PASS
	E6-288 primer&probe MIX		PASS
	E7-110 primer&probe MIX		PASS
	E7-240 primer&probe MIX		PASS
No Template Control (NTC)	The detection value of NTC should be 2 larger than the Ct value of the lowest standard curve concentration.		PASS

Approver Zong WeiyiDAY/MONTH/YEAR 12/08/2025
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China