

Document No.: Human-BRP-ZL-03

Version No.: A/3

## Residual Human DNA Quantitation Kit

Product No.: SK030207H100

Lot No.: 20250601

Date of Mfg.: 30 Jun. 2025

Expiration Date: 11 Nov. 2026

Test Item	Specification	Result
Linearity of Standard curve	$R^2 \geq 0.990$ .	PASS
Amplification efficiency of Standard curve	$E=90.0\%-110.0\%$ .	PASS
Repeatability	30 fg/ $\mu$ L Human DNA Control was detected, $CV \leq 15.0\%$ .	PASS
No Template Control (NTC)	The detection value of NTC should not be more than 6.00 fg/ $\mu$ L.	PASS

Approved By Zong WeiyingDAY/MONTH/YEAR 02/07/2025  
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access [www.shentekbio.com](http://www.shentekbio.com)For any other inquiries, please contact us by email: [info@shentekbio.com](mailto:info@shentekbio.com)

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China