

Document No.: NS0-BRP-ZL-04

Version No.: A/1

**Residual NS0&SP2/0 DNA Quantitation Kit**Product No.: **SK030208N100**Lot No.: **20250601**Date of Mfg.: **11 Jun. 2025**Expiration Date: **14 Oct. 2026**

Test Item	Specification	Result
Linearity of Standard curve	$R^2 \geq 0.990$ .	PASS
Amplification efficiency of Standard curve	$E=90.0\%-110.0\%$ .	PASS
Repeatability	3 fg/ $\mu$ L NS0&SP2/0 DNA Control was detected, $CV \leq 20.0\%$ .	PASS
No Template Control (NTC)	The Ct value of NTC should not be less than 35.00 cycles or undetermined.	PASS

Approved By Zhang LipingDAY/MONTH/YEAR 16/06/2025  
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access [www.shentekbio.com](http://www.shentekbio.com)For any other inquiries, please contact us by email: [info@shentekbio.com](mailto:info@shentekbio.com)

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China