

Certificate of Analysis

Document No.: Pichia-BRP-ZL-01

Version No.: A/0

Residual Pichia Pastoris DNA Quantitation Kit

Product No.: SK030205P100

Lot No.: 20250201

Date of Mfg.: 10 Feb. 2025

Expiration Date: 26 Nov. 2026

Test Item	Specification	Result
Linearity of Standard curve	$R^2 \ge 0.990$.	PASS
Amplification efficiency of Standard curve	E=90.0%-110.0%.	PASS
Repeatability	30 fg/µL <i>Pichia Pastoris</i> DNA Control was detected, CV≤15.0%.	PASS
No Template Control (NTC)	The Ct value of NTC should be not less than 35.00 cycles or undetermined.	PASS

Approved By Zong Weight

DAY/MONTH/YEAR 19/02/WW.
Huzhou Shenke Biotechnology Co., Ltd.

Certificates available online, please access www.shentekbio.com

For any other inquiries, please contact us by email: info@shentekbio.com

Address:8th Floor,6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China