

## **Certificate of Analysis**

Document No.: Pichia-BRP-ZL-01

Version No.: A/1

## Residual Pichia Pastoris DNA Quantitation Kit

Product No.: **SK030205P100** 

Lot No.: 20250301

Date of Mfg.: 19 Mar. 2025

Expiration Date: 26 Nov. 2026

Test Item	Specification	Result
Linearity of Standard curve	$R^2 \ge 0.990.$	PASS
Amplification efficiency of Standard curve	E=90.0%-110.0%.	PASS
Repeatability	30 fg/µL <i>Pichia Pastoris</i> DNA Control was detected, CV≤15.0%.	PASS
No Template Control (NTC)	The Ct value of NTC should not be less than 35.00 cycles or undetermined.	PASS

Approved By Zong Weinig

Certificates available online, please access www.shentekbio.com
For any other inquiries, please contact us by email: info@shentekbio.com
Address:8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China