

Document No.: Plasm-BRP-ZL-07

Version No.: A/2

**Residual Plasmid DNA Quantitation Kit (3G)**

Product No.: 1101111-1

Lot No.: 20250401

Date of Mfg.: 07 Apr. 2025

Expiration Date: 03 Dec. 2025

Test Item	Specification	Result
Linearity of Standard curve	$R^2 \geq 0.990$ .	PASS
Amplification efficiency of Standard curve	$E = 90.0\% - 110.0\%$ .	PASS
Repeatability	Low concentration sample was detected, $CV \leq 20.0\%$ .	PASS
No Template Control (NTC)	The Ct value of NTC should be 2 larger than the Ct value of ST5.	PASS

Approved By Zong WeipingDAY/MONTH/YEAR 23/04/2025  
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access [www.shentekbio.com](http://www.shentekbio.com)For any other inquiries, please contact us by email: [info@shentekbio.com](mailto:info@shentekbio.com)

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China