

Document No.: SV40LTA & E1A-BRP-ZL-05

Version No.: A/0

Residual SV40LTA & E1A DNA Quantitation Kit (2G)Product No.: **1403443**Lot No.: **20250301**Date of Mfg.: **12 Mar. 2025**Expiration Date: **28 Feb. 2026**

Test Item	Specification	Result
Linearity of Standard curve	$R^2 \geq 0.990$.	PASS
Amplification efficiency of Standard curve	$E = 90.0\% - 110.0\%$.	PASS
Repeatability	Low concentration samples were detected, $CV \leq 20.0\%$.	PASS
No Template Control (NTC)	The SV40LTA detection value of NTC should be not more than 14.32 copies/ μ L, the E1A detection value of NTC should be not more than 17.79 copies/ μ L.	PASS

Approved By Zong WeijiangDAY/MONTH/YEAR 25/03/2025
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China