

Document No.: Vero-BRP-ZL-16

Version No.: A/0

**Residual Vero DNA-154 Quantitation Kit**Product No.: **SK030227V100**Lot No.: **20250702**Date of Mfg.: **28 Jul. 2025**Expiration Date: **05 Nov. 2026**

Test Item	Specification	Result
Linearity of Standard curve	$R^2 \geq 0.990$ .	PASS
Amplification efficiency of Standard curve	$E = 90.0\%-110.0\%$ .	PASS
Repeatability	3 fg/ $\mu$ L Vero DNA Control was detected, $CV \leq 20.0\%$ .	PASS
No Template Control (NTC)	The Ct value of NTC should not be less than 35.00 or undetermined.	PASS

Approved By Zong WeiyiDAY/MONTH/YEAR 04/08/2025  
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access [www.shentekbio.com](http://www.shentekbio.com)For any other inquiries, please contact us by email: [info@shentekbio.com](mailto:info@shentekbio.com)

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China