

Document No.: Vero-BRP-ZL-03

Version No.: A/1

Residual Vero DNA Quantitation KitProduct No.: **SK030204V100**Lot No.: **20250201**Date of Mfg.: **10 Feb. 2025**Expiration Date: **05 Nov. 2026**

Test Item	Specification	Result
Linearity of Standard curve	$R^2 \geq 0.990$.	PASS
Amplification efficiency of Standard curve	$E = 90.0\% - 110.0\%$.	PASS
Repeatability	3 fg/ μ L Vero DNA Control was detected, $CV \leq 15.0\%$.	PASS
No Template Control (NTC)	The Ct value of NTC should be not less than 35.00 cycles or undetermined.	PASS

Approved By Zong WeijiangDAY/MONTH/YEAR 17/02/2025
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China