

Document No.: Vero-BRP-ZL-03

Version No.: A/2

Residual Vero DNA Quantitation KitProduct No.: **SK030204V100**Lot No.: **20250601**Date of Mfg.: **11 Jun. 2025**Expiration Date: **09 Jan. 2027**

Test Item	Specification	Result
Linearity of Standard curve	$R^2 \geq 0.990$.	PASS
Amplification efficiency of Standard curve	$E=90.0\%-110.0\%$.	PASS
Repeatability	3 fg/ μ L Vero DNA Control was detected, $CV \leq 15.0\%$.	PASS
No Template Control (NTC)	The Ct value of NTC should not be less than 35.00 cycles or undetermined.	PASS

Approved By Zhang LipingDAY/MONTH/YEAR 16/06/2025
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China