

Certificate of Analysis

Document No.: Vero-BRP-ZL-10

Version No.: A/2

Residual Vero DNA Size Analysis Kit (2G)

Product No.: 1103174

Lot No.: 20250401

Date of Mfg.: 11 Apr. 2025

Expiration Date: 05 Nov. 2026

Test Item	Specification		Result
Linearity of Standard curve	R ² ≥0.990.		PASS
Amplification efficiency of Standard curve	E=83.3%-110.0%.		PASS
Repeatability	Vero Primer&Probe MIX-85	30 fg/μL Vero DNA Control was detected, CV≤20.0%.	PASS
	Vero Primer&Probe MIX-134		PASS
	Vero Primer&Probe MIX-229		PASS
	Vero Primer&Probe MIX-552		PASS
No Template Control (NTC)	The Ct value of NTC should be 2 larger than the Ct value of ST5.		PASS

Approved By Zong Weining

DAY/MONTH/YEAR // O4/2025
Huzhou Shenke Biotechnology Co., Ltd.

Certificates available online, please access www.shentekbio.com

For any other inquiries, please contact us by email: info@shentekbio.com

Address:8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China