

Document No.: Reverse Transcriptase-BRP-ZL-02

Version No.: A/1

Reverse Transcriptase Assay ControlProduct No.: **1505701**Lot No.: **20250301**Date of Mfg.: **04 Mar. 2025**Expiration Date: **30 Oct. 2025**

| Test Item | Specification | Result |
|-------------------------------|---|--------|
| Positive Control Sample (PCS) | The Ct value of PCS should be not more than 28.00. | PASS |
| Negative Control Sample (NCS) | The Ct value of NCS (Buffer A) should be not less than 40.00. | PASS |

Approved By Zong WeijiangDAY/MONTH/YEAR 13/03/2025
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China