

Document No.: Virus-BRP-ZL-46

Version No.: A/0

Vesivirus 2117 Detection KitProduct No.: **1506743**Lot No.: **20250301**Date of Mfg.: **20 Mar. 2025**Expiration Date: **23 Dec. 2026**

Test Item	Specification	Result
Low concentration sample	FAM channel: Ct < 40.00 with specific amplification. VIC channel: Ct < 35.00 with specific amplification.	PASS
No Template Control (NTC)	FAM channel: Undetected or absence of specific amplification. VIC channel: Ct < 35.00 with specific amplification.	PASS

Approved By Zong WeijieDAY/MONTH/YEAR 27/03/2025
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China