

Document No.: Virus-BRP-ZL-48

Version No.: A/0

Vesivirus 2117 Positive Control

Product No.: 1506743-R01

Lot No.: 20250301

Date of Mfg.: 26 Mar. 2025

Expiration Date: 11 Feb. 2026

Test Item	Specification	Result
Positive Control Sample (PCS)	FAM channel: Ct < 35.00 with specific amplification. VIC channel: Ct < 35.00 with specific amplification.	PASS
Negative Control Sample (NCS)	FAM channel: Undetected or absence of specific amplification. VIC channel: Ct < 35.00 with specific amplification.	PASS

Approved By Zong WeiyangDAY/MONTH/YEAR 28/03/2025
Huzhou Shenke Biotechnology Co., Ltd.Certificates available online, please access www.shentekbio.comFor any other inquiries, please contact us by email: info@shentekbio.com

Address: 8th Floor, 6B Building, No. 1366 Hongfeng Road, Huzhou 313000, Zhejiang Province, China